

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product	General Medical Devices Sanitary aids
Name and address of the applicant	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2 62-052 Komorniki, POLAND
Name and address of the manufacturer	ArjoHuntleigh AB Hans Michelsensgatan 10, 211 20 Malmö, SWEDEN
Name and address of the factory	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2, 62-052 Komorniki, POLAND
Ratings and principal characteristics	Rated voltage 24 Vdc Rated power 150 VA Degree of protection IP24 Type of applied part BF Operating time 10% (1 min. ON / 9 min. OFF) Max. patient weight 140 kg / 309 lb Max. dimensions 1310 x 624 x 1555 mm
Trade mark (if any)	ArjoHuntleigh
Model/type Ref.	Carino
Additional information (if necessary)	Certificate DE 3 - 40602 issued 2016-09-23 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015
as shown in the Test Report Ref. No. which forms part of this certificate	713110420

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-10-30  
CB 17 09 65428 013



Stefan Reis



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service