



Ref. Certif. No.

DE 3 - 40916

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product	Sanitary aids Shower trolley
Name and address of the applicant	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2 62-052 Komorniki, POLAND
Name and address of the manufacturer	ArjoHuntleigh AB Hans Michelsensgatan 10, 211 20 Malmö, SWEDEN
Name and address of the factory	ArjoHuntleigh Polska Sp. z o.o. ul. Ks. Piotra Wawrzyniaka 2, 62-052 Komorniki, POLAND
Ratings and principal characteristics	Rated voltage: 24 Vdc Rated power: Max. 216 VA Degree of protection: IP24 Type of applied part: BF Operating time: 10% (2 min. ON / 18 min. OFF) Max. patient weight: 182 kg / 400 lbs Safe working load: 200 kg / 440 lbs Max. dimensions (LxWxH): 2005 x 888 x 1315 mm
Trade mark (if any)	ArjoHuntleigh
Model/type Ref.	Carevo
Additional information (if necessary)	Certificate DE 3 - 40039 issued 2014-02-04 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015
as shown in the Test Report Ref. No. which forms part of this certificate	713110425

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2018-03-14
CB 18 02 65428 018

Manfred Appel



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service