



## Serviceanvisning LENA rollator med albuestøtte, varenummer: TGR-R RS 8811

### 6. Maintenance

- Clean the frame with damp cloth and dry thoroughly.
- Do not use strong polishing agents due to the damage of protecting coatings.
- Clean the impurities from moving parts of the rollator regularly (hinges, clamping screws, locks).

### 7. Storage and transport conditions

Device should not be exposed to direct sunlight, water or moisture.

