

CERTIFICATE

No. Z1 15 02 42097 059

Holder of Certificate: Invacare Rea AB

Org. No. 556098-9211

Växjövägen 303 343 71 Diö **SWEDEN**

Production Facility(ies):

Product:

42097

Certification Mark:



Medical Beds

Model(s): **SB755**

Parameters: Voltage: Frequency:

> Current: max. 1.5 A Degree of protection: IPX4 / IPX6 / IPX6W

230 VAC

50 Hz

Protection class: Operating time (INT): 10% (2 min/18min) Safety working load: 235 kg

Max. patient weight: 200 kg Applied part: Type B Application environment: 3 & 4

Dimension (width x length): 100 x 213 cm Mattress support platform: 90 x 200 cm

35 - 82 cm Bed to floor height: Mattress support platform: 4-sectioned

Tested according to: EN 60601-1:2006

EN 60601-2-52:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713055401 Valid until: 2020-02-02

Date, 2015-02-03 (Torsten Zimmer)

Page 1 of 1

