




TEST REPORT

EN 12184

Electrically powered wheelchairs, scooters and their chargers

— Requirements and test methods

Report reference No.	SHES230801672001-M2
Date of issue	2023-12-12
	Modification 1: 2025-04-25, Modification 2: 2025-09-30
Test by (name + signature)	Jack Lv 
Approved by (name + signature) :	Jason Gong 
Testing laboratory	SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.
Address	No. 588 West Jindu Rd, Xinqiao Town, Songjiang District 201612 Shanghai China.
Applicant	Anhui JBH Medical Apparatus Co., Ltd.
Address	No. 116 Qicang Road, Industrial Park, Mingguang, Anhui China.
Test Specification:	
Standard	EN 12184:2022
Test procedure	Test Report
Non-standard test method	N/A
Test item description	Powered Wheelchair
Trademark	
Manufacturer	Anhui JBH Medical Apparatus Co., Ltd.
Model/type reference	D26
Series/Batch No.	D26202503001
Type Class& Maximum Speed	Class B, 6km/h
Maximum Occupant mass	180 kg
Power Rating	Rechargeable Li-ion battery (Model: MFD16): 24 VDC, 9 Ah x 2, 216 Wh x 2 DC Motor (Model: MF003-250): 250 W x 2 Battery charger (Model: ZB2402ML) Input: 100-240 VAC, 50/60Hz, 1.5A Output: 24Vdc, 2A