



Human Handling Concept

CE

EU DECLARATION OF CONFORMITY

Name and address of the firm:

master care A/S Sofienlystvej 3, 8340 Malling, Denmark

Single Registration Number (SRN):

DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification:

Medical Device, Classification 1.

Slide Master, Unique Slide, Master Slide CT

Product name:

57440019763291QW

Basic-UDI-DI:

Article no.: MC 002-1240, MC 002-1242*, MC 002-1243, MC 002-1245*, MC 002-1246, MC 002-1277*, MC 002-1340, MC 002-1345, MC 002-1346, MC 002-1385, MC 002-1400, MC 002-1471, MC 002-1472, MC 002-1473, MC 002-1475, MC 002-1477. (*covers for main product)

Intended use: Patted Sliding Tube which reduces the friction under a user with reduced function during turning and transfers in bed.

Declaration's validity: 5 years after signing or changes to the technical documentation.

> master care The Boy Harris Stripping the Boy Barrier Care The Boy Harris Indentistancere.dk

Date of Issue and place: Malling, 15.01.2025 Name and function:

