

DECLARATION OF CONFORMITY

This certificate is to declare the conformance to the
Radio Equipment Directive 2014/53/EU for the manufacturing firm of

*Starkey Laboratories, Inc.,
6700 Washington Avenue South,
Eden Prairie, MN., USA 55344*
doing business as
Starkey Hearing Technologies

with confirmation from the European Representative of

*STARKEY Laboratories (Germany) G.m.b.H
Weg beim Jäger 218-222
22335 Hamburg*

Starkey hearing aid accessory entitled

Starlink Edge Remote Microphone Model 960

Starkey Laboratories hereby declares that the equipment specified above conforms to the aforementioned
directive and is CE-marked accordingly:

Article 3.1a: *Standard(s)* used to show compliance with protection of health and safety:

RF Exposure (Specific Absorption Rate (SAR)) US FCC Part 2.1091 limits, which are more stringent than
European ICNIRP requirements.

Compliant test report: Element STAK0353.5

Safety (Medical Device): IEC 60601-1, edition 3.2 and IEC 60601-1-11, edition 2.0.

Compliant test reports: Intertek 105939097MIN-001 and 011.

Safety (EMC): IEC 60601-1-2, edition 4.1.

Compliant test report: Element STAK0353.3

Article 3.1b: *Standard(s)* used to show compliance with EMC Directive, 2014/30/EU:

EN 301-489-17, v 3.3.1.

Compliant test report: Element STAK0353.4

Article 3.2: *Standard* used to show compliance:

EN 300 328, v 2.2.2.

Compliant test reports: Element STAK0353.1 and Starkey report GQ-RPT-10969

Conformity was assessed via Annex II, CONFORMITY ASSESSMENT PROCEDURE REFERRED TO IN
ARTICLE 17



Signature

William J. Mitchell

Full Name

Principal Wireless and Safety Regulatory Engineer

Position

Starkey Laboratories, Inc.

Firm

28 January 2025

Date