

We, ViewPlus Technologies, Inc.

1965 SW Airport Ave.

Corvallis, Oregon 97333, USA

declare under our sole responsibility that our products listed below meet these directives and standards:

Authorized European Representative

ViewPlus GmbH

Kolpingweg 15

49492 Westerkappeln, Germany

+49 (0) 5404 – 9596 414

e-mail: info@viewplus.com

Application of Council Directive(s):

EMC directive 89/336/EECAmending

Directives: 92/31/EEC,93/68/EEC, 98/13/EEC

Medical Device Regulation: (EU) 2017/745

Low Voltage Directive 2006/95/EC

Test Description	Specification	Test Method
Radiated Emissions	EN 55022: 2006 Class B	CISPR 22:2005 (Amended by A1:2005 and A2:2006)
Conducted Emissions	EN 55022: 2006 Class B	CISPR 22:2005 (Amended by A1:2005 and A2:2006)
Flicker	EN 61000-3-3:1995 (Amended by A1:2001 and A2:2005)	IEC 61000-3-3:2005
Harmonics	EN 61000-3-2:2006	IEC 61000-3-2:2005
Radiated Emissions	CISPR 22:2005 (Amended by A1:2005 and A2:2006) Class B	CISPR 22:2005 (Amended by A1:2005 and A2:2006)
Conducted Emissions	FCC 15.107:2007	ANSI C63.4:2003
ESD	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-2:2001
Radiated Immunity	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-3:2006 (Amended by A1:2007)
EFT	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-4:2004
Surge	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-5:2005
Conducted Immunity	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-6:2006
Magnetic Field Immunity	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-8:2001
Voltage Dips and Interrupts	EN 55024:1998 (Amended by A1:2001 and A2:2003)	IEC 61000-4-11:2004
EN12182:2012	Technical aids for disabled persons	IEC 60950-1

During the conformity assessment procedure of the Medical Device Directive, a risk analysis in accordance to EN ISO14971 has been performed for the products listed below.

Class 1 according to annex VIII of MDR (EU) 2017/745

Type of Equipment:Information Technology Equipment for Office/Home Office Use

Class B

Model No. Model 080802 Emprint

Year of Manufacture: from 2006

Model No. Models 2212XX Elite 200, 1012XX Premier 100

Model No. 40803-XX Cub Jr, 080803-XX Cub, 081403-XX Max

Year of Manufacture: from 2009

Model No. Model 241201 EmFuse

Year of Manufacture: from 2011

Additional for product below following test procedure has been performed.

Test Description	Specification	Test Method
An introduction to the New Safety Standard for ICT and AV Equipment	Standard: IEC 62368-1:2014 Test procedure: C8 Scheme Non-standard test method: N/A	IEC 62368-1:2014 (Second Edition)


Model No. Model VPTT1702-XX (XX = 00 or 01) EmBraille

Year of Manufacture: from 2019

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

Place: Corvallis, Oregon USA

Date: 17 September 2020



Carolyn K Gardner
Corporate Secretary